

November 3, 2025

To All Bidders:

Subject: Bid No. CR-8587-J6 – Maintenance, Service and Installation – AED's / WPS

ADDENDUM NO. 1

To Whom It May Concern:

With reference to our bid request relative to the above subject, please refer to the changes/modifications/clarifications to the original proposal request.

General Bid Clarification

• Please see the attached for the revised specifications page to be provided with your bid submission.

Bidders are requested to acknowledge and/or include this addendum with submission. All other terms, conditions and specifications remain unchanged.

Very truly yours,

Jerry Kucera Buyer

Bid #: CR-8587-J6 - Specifications (Addendum 1) Maintenance, Repair & Installation – AED's / WPS

Purchase of new AEDs: Any new units purchased shall be Powerheart G5 with ICPR & Sleeve Duo package

Each unit must include:

- Automatic with ICPR and sleeve
- Fully automatic with ICPR, dual language (English and Latin American Spanish)
- Battery
- 1 set of feedback defibrillation pads
- Carry sleeve
- AED manager
- USB cable
- 1 set user guide with step-by-step rescue guide
- Multilingual Electrode instructions for use
- Adult pads
- Pediatric pads
- 2-hour training upon request
- Must replace old AEDs with new units
- Old units must be removed and return must provide rebates
- Installation of new units. (Installation sites will be provided upon award.)

Service Plan for AED's in all buildings

330 Pluscare Advanced 2 RRS G5 plan must include:

- 2 years of coverage
- Bi-Annual service of each AED
- Certified technician will perform ON SITE bi-annual recommended maintenance and will replace any expired or soon to expire consumables
- Contract will cover all consumables
- All work must be documented and a copy of all documentation shall be forwarded to *Director of Nursing*.
- 3 AED training units
- Locations of all AED's will be provided upon award